

P	Δ	Δ	3	R	3	N	h	ı	n	1

Polyclonal Antibody to Phosphoinositide-3-Kinase Class-2-Alpha Polypeptide (PIK3C2a)

Organism Species: Mus musculus (Mouse)

Instruction manual

FOR RESEARCH USE ONLY
NOT FOR USE IN CLINICAL DIAGNOSTIC PROCEDURES

13th Edition (Revised in Aug, 2023)



[PROPERTIES]

Source: Polyclonal antibody preparation

Host: Rabbit

Purification: Antigen-specific affinity chromatography followed by Protein A affinity

chromatography

Traits: Liquid

Concentration: 0.5mg/mL

UOM: 100µL

Cross Reactivity: Human;Rat

Applications: WB; IHC; ICC; IP.

[IMMUNOGEN]

Immunogen: Recombinant PIK3C2a (Lys349~Pro611 (Accession # Q61194)) expressed in

E.coli

Accession No.: RPA383Mu01

[APPLICATIONS]

Western blotting: 0.01-2µg/mL;

Immunohistochemistry: 5-20µg/mL;

Immunocytochemistry: 5-20µg/mL;

Optimal working dilutions must be determined by end user.

[FORMULATION]

glycerol.

Form & Buffer: Supplied as solution form in PBS, pH7.4, containing 0.02% NaN3, 50%

[STORAGE AND STABILITY]

Storage: Avoid repeated freeze/thaw cycles.

Store at 4°C for frequent use.


Aliquot and store at -20°C for 24 months.


Stability Test: The thermal stability is described by the loss rate. The loss rate was determined

Cloud-Clone Corp.

by accelerated thermal degradation test, that is, incubate the protein at 37°C for 48h, and no obvious degradation and precipitation were observed. The loss rate is less than 5% within the expiration date under appropriate storage condition.

[IDENTIFICATION]





DAB staining on IHC-P; Samples:
Mouse Kidney Tissue; Primary Ab:
10µg/ml Rabbit Anti-Mouse PIK3C2a
Antibody Second Ab: 2µg/mL HRPLinked Caprine Anti-Rabbit IgG
Polyclonal Antibody (Catalog:

SAA544Rb19)

DAB staining on IHC-P;
Samples: Mouse Stomach Tissue;
Primary Ab: 10µg/ml Rabbit Anti-Mouse
PIK3C2a Antibody
Second Ab: 2µg/mL HRP-Linked
Caprine Anti-Rabbit IgG Polyclonal
Antibody
(Catalog: SAA544Rb19)

[IMPORTANT NOTE]

The kit is designed for research use only, we will not be responsible for any issue if the kit was used in clinical diagnostic or any other procedures.