

RPC488Hu01 50µg

**Recombinant Ferrochelatase (FECH)** 

Organism Species: Homo sapiens (Human)

Instruction manual

FOR RESEARCH USE ONLY NOT FOR USE IN CLINICAL DIAGNOSTIC PROCEDURES

13th Edition (Revised in Aug, 2023)



## [ PROPERTIES ]

**Source:** Prokaryotic expression

Host: E.coli

Residues: Gly55~Leu423

Tags: N-terminal His Tag

**Subcellular Location:** Mitochondrion

**Purity:** > 80%

Traits: Freeze-dried powder

**Buffer formulation:** PBS, pH7.4, containing 0.01% SKL, 5% Trehalose .

Original Concentration: 400µg/mL

Applications: Positive Control; Immunogen; SDS-PAGE; WB.

(May be suitable for use in other assays to be determined by the end user.)

Predicted isoelectric point: 8.6

Predicted Molecular Mass: 45.9kDa

**Accurate Molecular Mass:** 43kDa as determined by SDS-PAGE reducing conditions.

#### [USAGE]

Reconstitute in 10mM PBS (pH7.4) to a concentration of 0.1-1.0 mg/mL. Do not vortex.

#### [STORAGE AND STABILITY]

Storage: Avoid repeated freeze/thaw cycles.

Store at 2-8°C for one month.

Aliquot and store at -80°C for 12 months.

**Stability Test:** The thermal stability is described by the loss rate. The loss rate was determined by accelerated thermal degradation test, that is, incubate the protein at 37°C for 48h, and no obvious degradation and precipitation were observed. The loss rate is less than 5% within the expiration date under appropriate storage condition.

## [ SEQUENCE ]

```
GAKPQV QPQKRKPKTG ILMLNMGGPE TLGDVHDFLL RLFLDRDLMT LPIQNKLAPF IAKRRTPKIQ EQYRRIGGGS PIKIWTSKQG EGMVKLLDEL SPNTAPHKYY IGFRYVHPLT EEAIEEMERD GLERAIAFTQ YPQYSCSTTG SSLNAIYRYY NQVGRKPTMK WSTIDRWPTH HLLIQCFADH ILKELDHFPL EKRSEVVILF SAHSLPMSVV NRGDPYPQEV SATVQKVMER LEYCNPYRLV WQSKVGPMPW LGPQTDESIK GLCERGRKNI LLVPIAFTSD HIETLYELDI EYSQVLAKEC GVENIRRAES LNGNPLFSKA LADLVHSHIQ SNELCSKQLT LSCPLCVNPV CRETKSFFTS QQL
```

### [ IDENTIFICATION ]




Figure . Gene Sequencing (extract)




Figure. SDS-PAGE

# [ IMPORTANT NOTE ]

The kit is designed for research use only, we will not be responsible for any issue if the kit was used in clinical diagnostic or any other procedures.